

Title: WREMAC Cardiac Monitor/Defibrillator Specifications for All New Equipment Purchases	Effective Date: January 4, 2008
Policy # 2008-1	Page 1 of 2
	Revised:

POLICY:	Automated	BLSFR/EMT-B	EMT-I	EMT-CC/P
	Voice Prompts	Yes	Yes	Optional
	Visual Prompts	Yes	Yes	Optional
	Hands-Free Defibrillation	Optional	Optional	Optional
	Voice Recorder	Yes	Yes	Recommended
	Ability to Print Code Summary for Receiving Hospital Within 24 Hours	Yes	Yes	Yes
	Ability to Print Real-time Rhythm Strip	Optional (Device option may be available to CC/P and credentialed I's only)	Optional (Device option may be available to CC/P and credentialed I's only)	Yes
	Screen/Display to Monitor Rhythm	Optional (Device option may be available to CC/P and credentialed I's only)	Optional (Device option may be available to CC/P and credentialed I's only)	Yes
	Manual Operation (Adult & Pediatric)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Yes
	Synchronized Cardioversion (Adult & Pediatric)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Yes
	Pacing (Adult & Pediatric)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Yes
	Defibrillation	Yes	Yes	Yes

	(Adult & Pediatric)			
	Bi-Phasic Capabililty	Yes	Yes	Yes
	Waveform Capnography (Adult & Pediatric)	Optional (Device option may be available to I/CC/P only)	Yes	Yes
	12 Lead Monitoring Capability	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Optional (Device option may be available to CC/P only and controlled by lockout/password)	Yes
	Ability to transmit 12-lead EKG	N/A	N/A	Recommended
Procedure				
Reference				

YES- Must have feature
NO – Cannot have feature
Recommended – Should have feature
Optional – May have feature
N/A – Not applicable